

Ordering Data

Items supplied:

Each Transferpette® *S* adjustable and fixed volume pipette has DE-M marking and is supplied with performance certificate, shelf/rack mount and silicone oil.

Transferpette® *S*, adjustable volume

Capacity, µl (color-coded)	Description	A* ≤ ± %	A* ≤ ± µl	CV* ≤ %	CV* ≤ µl	Subdivision µl	Cat. No.
0.1 - 1	D-1	2	0.02	1.2	0.012	0.001	**7047 68**
0.1 - 2.5	D-2.5	1.4	0.035	0.7	0.018	0.002	**7047 69**
0.5 - 10	D-10	1	0.1	0.5	0.05	0.01	**7047 70**
2 - 20	D-20**	0.8	0.16	0.4	0.08	0.02	**7047 72**
5 - 50	D-50	0.8	0.4	0.3	0.15	0.05	**7047 73**
10 - 100	D-100	0.6	0.6	0.2	0.2	0.1	**7047 74**
20 - 200	D-200**	0.6	1.2	0.2	0.4	0.2	**7047 78**
100 - 1000	D-1000	0.6	6	0.2	2	1	**7047 80**
500 - 5000	D-5000	0.6	30	0.2	10	5	**7047 82**
1000 - 10000	D-10000	0.6	60	0.2	20	10	**7047 84**



Transferpette® *S*, fixed volume

Capacity, µl (color-coded)	Description	A* ≤ ± %	A* ≤ ± µl	CV* ≤ %	CV* ≤ µl	Cat. No.
10	F-10	1	0.1	0.5	0.05	**7047 08**
20	F-20**	0.8	0.16	0.4	0.08	**7047 16**
25	F-25	0.8	0.2	0.4	0.1	**7047 20**
50	F-50	0.8	0.4	0.4	0.2	**7047 28**
100	F-100	0.6	0.6	0.2	0.2	**7047 38**
200	F-200**	0.6	1.2	0.2	0.4	**7047 44**
500	F-500	0.6	3	0.2	1	**7047 54**
1000	F-1000	0.6	6	0.2	2	**7047 62**

* Calibrated to deliver (TD, Ex). Error limits according to the nominal capacity (= maximum volume) indicated on the instrument, obtained with instrument and distilled water at equilibrium with ambient temperature at 20 °C, and with smooth, steady operation. The error limits are well within the limits of DIN EN ISO 8655-2. DE-M marking.
A = Accuracy, CV = Coefficient of variation

** For use only with 2-200 µl pipette tips



Accessories

(Other accessories for Transferpette® *S* pipettes can be found on page 52 and in the operating manual.)

Starter Kit

Items supplied: 3 Transferpette® *S* adjustable volume pipettes, 3 x TipBox (filled), 3 shelf/rack mounts.

Type	The kit includes the following Transferpette® *S* models	Cat. No.
MICRO	D-1, D-10, D-100	**7047 90**
MIDI	D-20, D-200, D-1000	**7047 91**
MACRO	D-1000, D-5000, D-10000	**7047 92**
STANDARD	D-10, D-100, D-1000	**7047 93**

Note! When ordering instruments with DAkkS calibration certificates, the prefix 'DAkkS' must be added to the order number, e.g., DAkkS 7047 28.

BRAND also offers calibration service at the factory lab. For more information, please see page 326.